



DIVISION OF PUBLIC HEALTH

Jim Doyle
Governor

1 WEST WILSON STREET
P O BOX 2659
MADISON WI 53701-2659

Karen E. Timberlake
Secretary

State of Wisconsin
Department of Health and Family Services

608-266-1568
FAX: 608-261-6392
dhfs.wisconsin.gov

Numbered Memo Series 09-06

June 2009

To: First Responder and Ambulance Service Providers
Service Medical Directors
EMS Training Centers
Hospitals

From: Brian Litza, Chief
Wisconsin Emergency Medical Services Section

Re: ERROR IDENTIFIED – BROSELOW TAPE 2007 EDITION A

The following information has been identified as an error in this product and should be corrected as soon as possible.

There is a typographical error on the **Broselow Tape 2007 Edition A**. The error only pertains to this edition of the Broselow Tape.

The error states that the glucagon dosage for children 3-4-5 kg thru white zones should be 0.5mg/kg/dose. The correct dosage for children weighing 3-5 kg is a dose of 0.5 mg glucagon. For children that fall within the blue through green zones the error states that the dosage should be 1 mg/kg/dose with maximum dose of 1 mg. The correct dosage for children in the blue through green zones is 1 mg of glucagon.

The following corrections should be made to any and all Broselow Tapes 2007 Edition A:

- Black out the mg/kg dose for both of these glucagon dosages on the tape
- The dosages for glucagon should read:
 - Glucagon 0.5 mg (3-4-5 kg through white zones)
 - Glucagon 1 mg (blue through green zones).

Again, the Broselow Tape 2007 Edition **B** does NOT have this error.